



AURELIA® TRANSFORM

3.2mil, Nitrile Powder Free Examination Gloves



Regulations	PPE Regulation (EU) 2016/425 Medical Device Regulation (EU) 2017/745, 93/42/EEC
Quality Standards	ISO 9001:2015 Quality Management Systems ISO 13485: 2016 Medical Devices Quality Management Systems
Harmonised Standards	BS EN ISO 374 - 1,2,4,5 Protective gloves against dangerous chemicals and microorganisms. BS EN 455 1-4 Medical gloves for single use. EN 420: 2003 + A1: 2009 Protective gloves. General requirements and test methods. BS EN 1184: 1997. Materials and articles in contact with foodstuffs.

AURELIA® GLOVES

A SUPERMAX HEALTHCARE BRAND

sales@supermax.co.uk

+441733 615215

Supermax Healthcare Ltd,

Unit 12-16 Titan Drive

Fengate, Peterborough

PE1 5XN

aureliagloves.co.uk



10 x 200


XS S M L XL

98895 98896 98897 98898 98899

Sizes & Dimensions

Single wall thickness

Physical Properties



XS 70+10 mm

S 80+10 mm

M 95+10 mm

L 110+10 mm

XL > 110 mm

min. 240 mm

min. 240 mm

min. 240 mm

min. 240 mm

min. 240 mm

① min. 0.10 mm

② min. 0.06 mm

③ min. 0.05 mm

Unaged

Aged

Tensile Strength

Min. 14.0 MPa

Min. 14.0 MPa

Elongation

Min. 500%

Min. 400%

Strength at Break

Min. 6N

Min. 6N



Nitrile



Powder Free



5 Years



Micro



3.6g +/- 0.2g



Ice Blue

